

EU & UK Declaration of Conformity

Mercado Medic AB
Tryffelslingan 14
SE-181 57 Lidingö, Sweden
EUDAMED SRN: SE-MF-000005191

hereby declare under our sole responsibility as a manufacturer that the product identified in section 1 *Product identification* below conform to the provisions of the EU regulations, EC directives, UK regulations and Swedish national legislations listed in section 2 *Regulatory framework* in this document.

1 Product identification

Product name: REAL® 9000 PLUS

Alternative names: REAL 9000 PLUS EL AUFSTEHHILFE

Basic UDI-DI: 732184D0019P

Risk class (Regulation (EU) 2017/745 Annex VIII): Class I

Intended use: The REAL® 9000 PLUS EL AUFSTEHHILFE device range consists of modular indoor work chairs designed to be used by people who need assistive devices to perform sitting dynamic activities, move by their own power or stand up from sitting. The REAL® 9000 PLUS EL AUFSTEHHILFE intends to make use of the user's physical ability and can therefore be adapted individually in a large number of designs. The device is also intended to be used by passive users who need to correct their seating position to varying degrees. In these cases, the device is configured to be moved and adjusted by an assistant. The device is designed to relieve muscles, joints, bones and relieve pains linked to passivity.

The REAL® 9000 PLUS EL AUFSTEHHILFE is designed and recommended for one or more of the following indications:

- Difficulty or inability to walk.
- Difficulty or inability to stand up from sitting.
- Difficulty in maintaining an adequate seating position.
- Pain or exhaustion as a result of everyday tasks at home or at work.
- Where use of a wheelchair is not suitable as a result of activities of daily living.

There are no known contraindications for use of the REAL® 9000 PLUS EL AUFSTEHHILFE. If the device is purchased without prescription from qualified healthcare professionals, the user should consult their doctor whether there are any contraindications.

2 Regulatory framework

The product identified in section 1 *Product identification* above conform to the provisions of the EU regulations, EC directives, UK regulations and Swedish national legislations below:

Regulation (EU) 2017/745 Medical Devices

Regulation (EC) No 1907/2006 Chemical Substances (REACH)

Directive 2011/65/EU Restriction of Hazardous Substances (RoHS)

Directive 2012/19/EU Waste Electrical and Electronic Equipment (WEEE)

Swedish legislation SFS 1993:584

Swedish legislation LVFS 2003:11

The Medical Devices Regulations 2002 (UK MDR 2002)

**The Restriction of the Use of Certain Hazardous Substances in Electrical
and Electronic Equipment Regulations 2012**

No common specifications or harmonised standards are published in the Official Journal of the European Union (OJEU) for Regulation (EU) 2017/745 on Medical Devices (MDR) on the date of issue of this document. The following standards are deemed appropriate and likely to be harmonised with MDR. They have been applied in the design and development of the product as a presumption of conformity.

EN ISO 21856:2022

IEC 60601-1:2005 + A1 + A2

ISO 20417 :2021

EN 10993-1:2009

EN 1335-1:2000

EN 1335-2:2018

EN ISO 14971:2020/A11:2021

EN 50581:2012

EN ISO 15223-1:2021

Signed for and on behalf of Mercado Medic AB.

Lidingö, Sweden

Date of issue: 2023-09-21



Andreas Teske
Managing Director